



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

1 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No               | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|--------------------|-----------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| Permanent Facility |                                                           |                                                                                                                             |                                                                              |                                                                                   |                                                  |
| 1                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current @ 50 Hz                                                                                                          | Using 6½ Digital multimeter by Direct method                                 | 100 mA to 10 A                                                                    | 0.2 % to 0.3 %                                   |
| 2                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current @ 50 Hz                                                                                                          | Using 6½ Digital multimeter by Direct method                                 | 30 µA to 100 mA                                                                   | 0.3 % to 0.2 %                                   |
| 3                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC High Current @ 50 Hz                                                                                                     | Using 6½ Digit Multimeter and Current Transformer (2000/5A) by Direct Method | 10 A to 2000 A                                                                    | 0.5 %                                            |
| 4                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Resistance @1 kHz                                                                                                        | Using LCR meter by direct method                                             | 1 ohm to 10 kohm                                                                  | 0.16 % to 0.11 %                                 |
| 5                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Voltage @ 50 Hz                                                                                                          | Using 6½ Digital multimeter by Direct method                                 | 1 mV to 500 mV                                                                    | 0.8 % to 0.2 %                                   |
| 6                  | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Voltage @ 50 Hz                                                                                                          | Using 6½ Digital multimeter by Direct method                                 | 500 mV to 1000 V                                                                  | 0.2 % to 0.14 %                                  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

2 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                         | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                          | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 7    | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure) | Capacitance @ 1 kHz                                                                                                         | Using LCR meter by direct method                                        | 10 pF to 1 µF                                                                     | 0.32 %                                           |
| 8    | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure) | Capacitance @ 100 Hz                                                                                                        | Using LCR meter by direct method                                        | 10 µF to 10 mF                                                                    | 0.4 %                                            |
| 9    | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure) | Energy - 3 Phase 4 wire (10 V to 440 V) (1A to 10A), Power Factor (0.1 to 1), (Lead/Lag)                                    | Using Three Phase Power Energy Logger / Calibrator by Comparison method | 0.3 kWh to 6 kWh                                                                  | 1.39 % to 1.43 %                                 |
| 10   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure) | Energy - Single Phase, 2 wire (10 V to 440 V) (1 A to 10 A), Power Factor (0.1 to 1), (Lead/ Lag)                           | Using Three Phase Power Energy logger/ Calibrator by Comparison method  | 0.3 kWh to 6 kWh                                                                  | 1.39 % to 1.43 %                                 |
| 11   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure) | Harmonic, Total Harmonic Distortion (THD) @ 50 Hz                                                                           | Using 3 Phase energy logger with Source by Direct / Comparison Method   | 30 V to 240 V (1 % to 10 %)                                                       | 0.7 %                                            |
| 12   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure) | Inductance @ 100 Hz                                                                                                         | Using LCR meter by direct method                                        | 0.1 mH to 10 H                                                                    | 0.24 %                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

3 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                    | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 13   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Current @ 50 Hz                                                                                                          | Using Multi Product Calibrator with current Coil by Direct method | 20 A to 1000 A                                                                    | 0.4 % to 1.5 %                                   |
| 14   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Current @ 50 Hz                                                                                                          | Using Multi Product Calibrator by Direct method                   | 3 mA to 20 A                                                                      | 0.12 % to 0.4 %                                  |
| 15   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Current @ 50 Hz                                                                                                          | Using Multi Product Calibrator by Direct method                   | 30 µA to 3 mA                                                                     | 0.55 % to 0.12 %                                 |
| 16   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Power (Active) 1 V to 1000V, 1 mA to 20A, UPF @ 50 Hz, 1 lead /Lag                                                       | Using Multi product Calibrator & Energy logger by Direct method   | 1 W to 12000 W                                                                    | 0.30 % to 0.09 %                                 |
| 17   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Voltage (@ 50 Hz)                                                                                                        | Using Multi Product Calibrator by Direct method                   | 10 mV to 100 V                                                                    | 0.35 % to 0.04 %                                 |
| 18   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Voltage (@ 50 Hz)                                                                                                        | Using Multi Product Calibrator by Direct method                   | 100 V to 1000 V                                                                   | 0.04 % to 0.06 %                                 |
| 19   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Voltage (50 HZ)                                                                                                          | Using Multi Product Calibrator by Direct method                   | 1 mV to 10 mV                                                                     | 2.5 % to 0.35 %                                  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

4 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 20   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Capacitance @ 1 kHz                                                                                                         | Using Multi Product Calibrator by Direct method                 | 0.5 nF to 10 nF                                                                   | 2.9 % to 0.58 %                                  |
| 21   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Capacitance @ 100 Hz                                                                                                        | Using Multi Product Calibrator by Direct method                 | 30 nF to 100 µF                                                                   | 0.5 % to 1.4 %                                   |
| 22   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Inductance @ 1 kHz                                                                                                          | Using LCR meter by direct method                                | 100 µH to 10 H                                                                    | 2.7 %                                            |
| 23   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Power Factor (Lead/Lag) 240V/ 5A @ 50 Hz                                                                                    | Using Multi product Calibrator & Energy logger by Direct method | 0.1 PF to 1 PF                                                                    | 0.054 PF                                         |
| 24   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)               | DC Current                                                                                                                  | Using 6½ Digital multimeter by Direct method                    | 10 µA to 100 mA                                                                   | 0.7 % to 0.3 %                                   |
| 25   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)               | DC Current                                                                                                                  | Using 6½ Digital multimeter by Direct method                    | 100 mA to 10 A                                                                    | 0.3 %                                            |
| 26   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)               | DC High Current                                                                                                             | Using 6½ Digit Multimeter and Current Shunt by VI Method        | 10 A to 100 A                                                                     | 0.75 %                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

5 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                         | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure           | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 27   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC High Current                                                                                                             | Using 6½ Digit Multimeter and Current shunt by VI Method | 100 A to 150 A                                                                    | 1.2 %                                            |
| 28   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC High Voltage                                                                                                             | Using HV probe with DMM by direct method                 | 1 kV to 35 kV                                                                     | 4.5 %                                            |
| 29   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC Voltage                                                                                                                  | Using 6½ Digital multimeter by Direct method             | 1 mV to 100 mV                                                                    | 0.4 % to 0.01 %                                  |
| 30   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC Voltage                                                                                                                  | Using 6½ Digital multimeter by Direct method             | 10 V to 1000 V                                                                    | 0.012 % to 0.007 %                               |
| 31   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC Voltage                                                                                                                  | Using 6½ Digital multimeter by Direct method             | 100 mV to 10 V                                                                    | 0.01 % to 0.012 %                                |
| 32   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance                                                                                                                  | Using Insulation Tester by direct method                 | 1 Mohm to 1 Tohm                                                                  | 2.5 % to 2.9 %                                   |
| 33   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance                                                                                                                  | Using 6½ Digital multimeter by Direct method             | 1 Mohm to 100 Mohm                                                                | 0.013 % to 0.96 %                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

6 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                         | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 34   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance                                                                                                                  | Using 6½ Digital multimeter by Direct method    | 1 Ohm to 1 Mohm                                                                   | 0.36 % to 0.013 %                                |
| 35   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance                                                                                                                  | Using 6½ Digital multimeter by Direct method    | 100 Mohm to 1000 Mohm                                                             | 0.96 % to 1.8 %                                  |
| 36   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance - 4 wire                                                                                                         | Using micro ohm meter by direct method          | 1 mohm to 100 mohm                                                                | 0.2 % to 0.05 %                                  |
| 37   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance - 4 wire                                                                                                         | Using microohm meter by direct method           | 10 µohm to 1 mohm                                                                 | 0.73 % to 0.32 %                                 |
| 38   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance - 4 Wire                                                                                                         | Using micro ohm meter by direct method          | 100 mohm to 1 ohm                                                                 | 0.05 % to 0.09 %                                 |
| 39   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)  | DC Current                                                                                                                  | Using Multi Product Calibrator by Direct method | 1 µA to 10 µA                                                                     | 2.50 % to 0.3 %                                  |
| 40   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)  | DC Current                                                                                                                  | Using Multi Product Calibrator by Direct method | 1 mA to 20 A                                                                      | 2.5 % to 0.4 %                                   |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

7 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 41   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Current                                                                                                                  | Using Multi Product Calibrator by Direct method                  | 10 $\mu$ A to 1 mA                                                                | 0.3 % to 2.5 %                                   |
| 42   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Current                                                                                                                  | Using Multiproduct Calibrator with Current Coil by Direct method | 20 A to 1000 A                                                                    | 0.4 % to 0.13 %                                  |
| 43   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Power (Active power) 10 V to 1000V, 0.1 A to 20A                                                                         | Using Multiproduct calibrator by direct method                   | 100 W to 20000 W                                                                  | 0.97 %                                           |
| 44   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Voltage                                                                                                                  | Using Multi Product Calibrator by Direct method                  | 1 mV to 330 mV                                                                    | 0.4 % to 0.015 %                                 |
| 45   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Voltage                                                                                                                  | Using Multi Product Calibrator by Direct method                  | 330 mV to 1000 V                                                                  | 0.015 % to 0.01 %                                |
| 46   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method                  | 1 Ohm to 30 Ohm                                                                   | 0.5 % to 0.07 %                                  |
| 47   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method                  | 30 kohm to 30 Mohm                                                                | 0.13 %                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

8 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 48   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method | 30 Mohm to 1000 Mohm                                                              | 0.13 % to 0.75 %                                 |
| 49   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method | 30 Ohm to 30 kohm                                                                 | 0.07 % to 0.13 %                                 |
| 50   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method      | 1 Tohm                                                                            | 3.5 %                                            |
| 51   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method      | 10 Gohm                                                                           | 3.5 %                                            |
| 52   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method      | 10 Mohm                                                                           | 3.5 %                                            |
| 53   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method      | 100 Gohm                                                                          | 3.5 %                                            |
| 54   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method      | 100 Mohm                                                                          | 3.7 %                                            |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

9 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 55   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method     | 200 Gohm                                                                          | 3.5 %                                            |
| 56   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 1 Tohm                                                                            | 3.5 %                                            |
| 57   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 10 Gohm                                                                           | 3.5 %                                            |
| 58   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 10 Mohm                                                                           | 3.5 %                                            |
| 59   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 100 Gohm                                                                          | 3.5 %                                            |
| 60   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 100 Mohm                                                                          | 3.5 %                                            |
| 61   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 200 Gohm                                                                          | 3.74 %                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

10 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 62   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 10 mohm                                                                           | 0.38 %                                           |
| 63   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 10 Ohm                                                                            | 0.2 %                                            |
| 64   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 100 mohm                                                                          | 0.2 %                                            |
| 65   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 wire                                                                                                         | Using Low Resistance Box by Direct method      | 1000 mohm                                                                         | 0.2 %                                            |
| 66   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 1 mohm                                                                            | 0.55 %                                           |
| 67   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 100 Ohm                                                                           | 0.25 %                                           |
| 68   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 1000 Ohm                                                                          | 0.25 %                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Validity** 25/06/2025 to 24/06/2029

**Page No** 11 of 81

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                              | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 69   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | ORP Meter / ORP Indicator                                                                                                   | Using Multi product calibrator By Direct Method                 | (-) 1500 mV to 1500 mV                                                            | 0.5 %                                            |
| 70   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | Oscilloscope Amplitude @ 1 kHz Sinewave                                                                                     | Using Function / Arbitrary wave form generator by Direct method | 2 mVpp to 20 Vpp                                                                  | 0.45 % to 0.35 %                                 |
| 71   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | Oscilloscope Amplitude @ 1 kHz Square wave                                                                                  | Using Function / Arbitrary wave form generator by Direct method | 2 mVpp to 20 Vpp                                                                  | 0.37 % to 0.35 %                                 |
| 72   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | Oscilloscope Bandwidth                                                                                                      | Using Function / Arbitrary wave form generator by Direct method | 50 kHz to 20 MHz                                                                  | 2.5 %                                            |
| 73   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | Oscilloscope Time Base                                                                                                      | Using Function / Arbitrary wave form generator by Direct method | 5 ns to 10 s                                                                      | 0.005 % to 0.01 %                                |
| 74   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | pH Meter (simulation method)                                                                                                | Using Multi product calibrator By Direct Method                 | 0 to 14 pH                                                                        | 0.13 %                                           |
| 75   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | TDS/ Conductivity Meter (Equivalent Resistance: 1 ohm to 2 Mohm)                                                            | Using Multiproduct calibrator by simulation method              | 0.5 $\mu$ S/cm to 1000 mS/cm                                                      | 1.72 % to 1.5 %                                  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

12 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 76   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | B Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 600 °C to 1800 °C                                                                 | 0.77 °C                                          |
| 77   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | E Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 250 °C to 1000 °C                                                             | 0.63 °C                                          |
| 78   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | J Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1200 °C                                                             | 0.40 °C                                          |
| 79   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | K Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1300 °C                                                             | 0.53 °C                                          |
| 80   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | N Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1300 °C                                                             | 0.54 °C                                          |
| 81   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | R Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 100 °C to 1700 °C                                                                 | 0.95 °C                                          |
| 82   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | RTD                                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 800 °C                                                              | 0.25 °C                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

13 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)( $\pm$ ) |
|------|----------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
| 83   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | S Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 100 °C to 1750 °C                                                                 | 0.65 °C                                                |
| 84   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | T Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 400 °C                                                              | 0.74 °C                                                |
| 85   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)  | B Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 600 °C to 1800 °C                                                                 | 0.68 °C                                                |
| 86   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)  | E Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 250 °C to 1000 °C                                                             | 0.60 °C                                                |
| 87   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)  | J Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1050 °C                                                             | 0.74 °C                                                |
| 88   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)  | K Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1350 °C                                                             | 0.47 °C                                                |
| 89   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)  | N Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1300 °C                                                             | 0.47 °C                                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

14 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 90   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | R Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method              | 0 °C to 1750 °C                                                                   | 0.67 °C                                          |
| 91   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | RTD                                                                                                                         | Using Multi Product Calibrator by Direct method              | (-) 200 °C to 800 °C                                                              | 0.3 °C                                           |
| 92   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | S Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method              | 0 °C to 1750 °C                                                                   | 0.66 °C                                          |
| 93   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | T Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method              | (-) 200 °C to 400 °C                                                              | 0.73 °C                                          |
| 94   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Frequency                                                                                                                   | Using 6½ Digital multimeter by Direct method                 | 10 Hz to 1 MHz                                                                    | 0.096 % to 0.02 %                                |
| 95   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Time                                                                                                                        | Using Digital Time Calibration standard by Comparison method | 0.1 s to 1 s                                                                      | 0.011 s                                          |
| 96   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Time                                                                                                                        | Using Digital Time Calibration standard by Comparison method | 1 s to 60 s                                                                       | 0.06 s                                           |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

15 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                           | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                          | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 97   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | Time                                                                                                                                                 | Using Digital Time Calibration standard by Comparison method | 3600 s to 43200 s                                                                 | 2.25 s                                           |
| 98   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | Time                                                                                                                                                 | Using Digital Time Calibration standard by Comparison method | 43200 s to 86400 s                                                                | 2.50 s                                           |
| 99   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure) | Time                                                                                                                                                 | Using Digital Time Calibration standard by Comparison method | 60 s to 3600 s                                                                    | 2.25 s                                           |
| 100  | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)  | Frequency                                                                                                                                            | Using Arbitrary Function generator by Direct method          | 1 MHz to 200 MHz                                                                  | 0.07 % to 0.5 %                                  |
| 101  | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)  | Frequency                                                                                                                                            | Using Multi Product Calibrator by Direct method              | 10 Hz to 2 MHz                                                                    | 0.5 % to 0.05 %                                  |
| 102  | MECHANICAL-ACCELERATION AND SPEED            | Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method           | 10 rpm to 100 rpm                                                                 | 0.46 rpm                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

16 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                          | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 103  | MECHANICAL-ACCELERATION AND SPEED | Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method                  | 10000 rpm to 30000 rpm                                                            | 6.01 rpm                                         |
| 104  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                | Using Standard digital tachometer & RPM source by Comparison method | 10 rpm to 100 rpm                                                                 | 0.46 rpm                                         |
| 105  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                | Using Standard digital tachometer & RPM source by Comparison method | 100 rpm to 1000 rpm                                                               | 0.65 rpm                                         |
| 106  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                | Using Standard digital tachometer & RPM source by Comparison method | 1000 rpm to 5000 rpm                                                              | 2.02 rpm                                         |
| 107  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                | Using Standard digital tachometer & RPM source by Comparison method | 10000 rpm to 30000 rpm                                                            | 6.01 rpm                                         |
| 108  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                | Using Standard digital tachometer & RPM source by Comparison method | 30000 rpm to 50000 rpm                                                            | 6.0 rpm                                          |
| 109  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                | Using Standard digital tachometer & RPM source by Comparison method | 5000 rpm to 10000 rpm                                                             | 2.02 rpm                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

17 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                       | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 110  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                             | Using Standard digital tachometer & RPM source by Comparison method | 50000 rpm to 70000 rpm                                                            | 7.01 rpm                                         |
| 111  | MECHANICAL-ACCELERATION AND SPEED | Digital Tachometer (Non-contact type)                                                                                                                             | Using Standard digital tachometer & RPM source by Comparison method | 70000 rpm to 90000 rpm                                                            | 9.2 rpm                                          |
| 112  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method                  | 1000 rpm to 5000 rpm                                                              | 2.02 rpm                                         |
| 113  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method                  | 30000 rpm to 50000 rpm                                                            | 6.0 rpm                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

18 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                       | Calibration or Measurement Method or procedure     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)( $\pm$ ) |
|------|-----------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
| 114  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 5000 rpm to 10000 rpm                                                             | 2.02 rpm                                               |
| 115  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 50000 rpm to 70000 rpm                                                            | 7.01 rpm                                               |
| 116  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 70000 rpm to 90000 rpm                                                            | 9.2 rpm                                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

19 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument   | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)( $\pm$ ) |
|------|-----------------------------------|-------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
| 117  | MECHANICAL-ACCELERATION AND SPEED | Indicator of RPM<br>Indicating device of Centrifuge, Bitumen extractor, Vibrating Machine, Abrasion Tester (Non-contact type) | Using Standard digital tachometer by direct method                  | 10 rpm to 100 rpm                                                                 | 0.46 rpm                                               |
| 118  | MECHANICAL-ACCELERATION AND SPEED | Indicator of RPM<br>Indicating device of Centrifuge, Bitumen extractor, Vibrating Machine, Abrasion Tester (Non-contact type) | Using Standard digital tachometer by direct method                  | 100 rpm to 1000 rpm                                                               | 0.65 rpm                                               |
| 119  | MECHANICAL-ACCELERATION AND SPEED | Indicator of RPM<br>Indicating device of Centrifuge, Bitumen extractor, Vibrating Machine, Abrasion Tester (Non-contact type) | Using Standard digital tachometer by direct method                  | 10000 rpm to 30000 rpm                                                            | 6.01 rpm                                               |
| 120  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                     | Using Standard digital tachometer & RPM source by Comparison method | 10 rpm to 100 rpm                                                                 | 0.45 rpm                                               |
| 121  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                     | Using Standard digital tachometer & RPM source by Comparison method | 100 rpm to 1000 rpm                                                               | 0.82 rpm                                               |
| 122  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                     | Using Standard digital tachometer & RPM source by Comparison method | 1000 rpm to 3000 rpm                                                              | 2.12 rpm                                               |
| 123  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                     | Using Standard digital tachometer & RPM source by Comparison method | 3000 rpm to 5000 rpm                                                              | 2.2 rpm                                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

20 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 124  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                                              | 10 rpm to 100 rpm                                                                 | 0.44 rpm                                         |
| 125  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                                              | 10 rpm to 100 rpm                                                                 | 0.45 rpm                                         |
| 126  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                                              | 100 rpm to 1000 rpm                                                               | 0.82 rpm                                         |
| 127  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                                              | 1000 rpm to 3000 rpm                                                              | 2.12 rpm                                         |
| 128  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer & RPM source by Comparison method                             | 3000 rpm to 5000 rpm                                                              | 2.2 rpm                                          |
| 129  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                                              | 3000 rpm to 5000 rpm                                                              | 2.2 rpm                                          |
| 130  | MECHANICAL-ACOUSTICS                                          | Sound Level Meter @ 1 kHz                                                                                                   | Using Sound Level Calibrator by Comparison Method                                               | 94 dB & 114 dB                                                                    | 0.99 dB                                          |
| 131  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Air Gauge Unit                                                                                                              | Using Air plug gauge & Air ring gauge by comparison method                                      | ± 100 µm                                                                          | 3.0 µm                                           |
| 132  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Angle Plate/Box angle plate /Try square Flatness of working face                                                            | Using Surface plate , Master cylinder , Lever Dial , Height gauge , Jacks- By Comparison Method | Upto 300 mm                                                                       | 13.0 µm                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

21 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                 | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 133  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Angle Plate/Box angle plate /Try square Squareness                                                                          | Using Surface plate , Master cylinder , Lever Dial , Height gauge , Jacks By Comparison Method | Upto 300 mm                                                                       | 13.0 µm                                          |
| 134  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Beam/Motar Mould (Dimension Inspection)                                                                                     | Using Digital Caliper by Direct Method                                                         | 0 to 600 mm                                                                       | 27.6 µm                                          |
| 135  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bench Centre (coaxility)                                                                                                    | Using Test mandrel & Dial test indicator by Comparison method                                  | Up to 300 mm                                                                      | 8.0 µm                                           |
| 136  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bench Centre (parallelism)                                                                                                  | Using Test mandrel & Dial test indicator by Comparison method                                  | Up to 300 mm                                                                      | 8.0 µm                                           |
| 137  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bevel Protractor/Combination set (L.C: 5 ')                                                                                 | Using Angle Gauge and Surface plate by Comparison method                                       | 0° to 360°                                                                        | 6.22 minute of arc                               |
| 138  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Box Angle Plate /angle plate/Try square Parallelism of opposite faces and edges                                             | Using Surface plate , Master cylinder , Lever Dial , Height gauge , Jack By Comparison Method: | Upto 300 mm                                                                       | 13.0 µm                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Validity** 25/06/2025 to 24/06/2029

**Page No** 22 of 81

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 139  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | CBR Mould - Diameter                                                                                                        | Using Vernier Caliper by Comparison Method     | 0.5 mm to 200 mm                                                                  | 21.0 µm                                          |
| 140  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | CBR Mould - Height                                                                                                          | Using Vernier Caliper by Comparison Method     | 0.5 mm to 300 mm                                                                  | 22.0 µm                                          |
| 141  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Chamfer /Weld fillet / Hi-Lo / Bridge cam gauge - Angle                                                                     | Using Profile projector by Comparison Method   | 1 ° to 90 °                                                                       | 10 minute of Arc                                 |
| 142  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Chamfer /Weld fillet / Hi-Lo / Bridge cam gauge- Length                                                                     | Using Profile projector by Comparison Method   | 0 to 60 mm                                                                        | 7.4 µm                                           |
| 143  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Coating Thickness Gauge (L.C: 0.01/0.1 µm)                                                                                  | Using Standard Foils by Comparison method      | 10 µm to 720 µm                                                                   | 2.3 µm                                           |
| 144  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Combination set, Clinometer , Inclinator L.C: 1°                                                                            | Using Angle Gauges- By Comparison Method       | 0 ° to 180 °                                                                      | 34.7 minute of arc                               |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

23 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 145  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Comparator Stand (Flatness of Base)                                                                                         | Using Dial Indicator by Comparison method                    | Up to 300 mm x 300 mm                                                             | 3.02 µm                                          |
| 146  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Core Cutter, Core Cutter Dolly - Diameter                                                                                   | Using Vernier Caliper by Comparison Method as per IS 2720-29 | 0 to 120 mm                                                                       | 24.9 µm                                          |
| 147  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Core Cutter, Core Cutter Dolly - Length<br>Core Cutter, Core Cutter Dolly - Length                                          | Using Vernier Caliper by Comparison Method as per IS 2720-29 | 0 to 150 mm                                                                       | 19.6 µm                                          |
| 148  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Core Cutter, Core Cutter Dolly - Thickness                                                                                  | Using Vernier Caliper by Comparison Method as per IS 2720-29 | 0.5 mm to 10 mm                                                                   | 17 µm                                            |
| 149  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cross Hatch cutter                                                                                                          | Using Profile Projector by Direct Method                     | 1 mm                                                                              | 6 µm                                             |
| 150  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cube Moulds                                                                                                                 | Using Digital Caliper by Comparison method                   | 20 mm to 200 mm                                                                   | 15 µm                                            |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

24 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                              | Calibration or Measurement Method or procedure             | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 151  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cylinder Mould - Diameter                                                                                                                                                | Using Vernier Caliper by Comparison Method as per IS 10086 | 0.5 mm to 160 mm                                                                  | 32.9 µm                                          |
| 152  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cylinder Mould - Height                                                                                                                                                  | Using Vernier Caliper by Comparison Method as per IS 10086 | 0.5 mm to 320 mm                                                                  | 17.1 µm                                          |
| 153  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cylinder Mould - Thickness                                                                                                                                               | Using Vernier Caliper by Comparison Method as per IS 10086 | 1 mm to 30 mm                                                                     | 22.9 µm                                          |
| 154  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cylindrical Mould (Diameter/Depth/Height) CBR Mould, Proctor Mould, Le-Chatelier Mould, Aggregate impact Crushing Mould, Compaction Mould, Vicate Mould, Density basket. | Using Digitak Caliper by Direct Method                     | 20 mm to 300 mm                                                                   | 37.2 µm                                          |
| 155  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Depth Caliper (L.C.: 0.01 mm and coarser)                                                                                                                                | Using Slip Gauge Set by Comparison Method                  | 0 to 300 mm                                                                       | 15.8 µm                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

25 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 156  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Depth Micrometer                                                                                                            | Using Slip gauge set by Comparison method          | 0 to 300 mm                                                                       | 4.3 µm                                           |
| 157  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Dial Bore Gauge (L.C: 0.01 mm)                                                                                              | Using Dial Calibration tester by Comparison method | 0 to 1 mm                                                                         | 6.5 µm                                           |
| 158  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Dial Thickness gauge (L.C: 0.001 mm)                                                                                        | Using Slip Gauge Block by Comparison method        | 0 to 10 mm                                                                        | 0.4 µm                                           |
| 159  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Dial Thickness gauge (L.C: 0.01 mm)                                                                                         | Using Slip Gauge Block by Comparison method        | 0 to 50 mm                                                                        | 5.8 µm                                           |
| 160  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Elongation Index/Flakiness Index apparatus                                                                                  | Using Digital Caliper by Comparison method         | 0 to 100 mm                                                                       | 25 µm                                            |
| 161  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer (L.C: 0.001 mm)                                                                                         | Using Slip Gauge Block by Comparison method        | 0 to 25 mm                                                                        | 0.76 µm                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

26 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                              | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 162  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer (L.C: 0.001 mm)                                                                                         | Using Slip gauge block by Comparison method                                 | 25 to 100 mm                                                                      | 1.2 µm                                           |
| 163  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer (L.C: 0.01 mm)                                                                                          | Using Slip Gauge Block by Comparison method                                 | 0 to 150 mm                                                                       | 6.0 µm                                           |
| 164  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer,Least count 0.01                                                                                        | Using gauge block and Long gauge block                                      | 100 mm to 600 mm                                                                  | 7.7 µm                                           |
| 165  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Feeler Gauge                                                                                                                | Using Digital Micrometer by Comparison method                               | 0.02 mm to 2 mm                                                                   | 3.4 µm                                           |
| 166  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Flush Pin Gauge                                                                                                             | Using Slip Gauge,Comparator Stand and electronic probe by comparison method | 4 mm to 150 mm                                                                    | 3.0 µm                                           |
| 167  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Foils                                                                                                                       | Using Electronic Probe & Comparator stand by Comparison method              | 0.01 mm to 1 mm                                                                   | 1.33 µm                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

27 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 168  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Hegman gauge depth measurement                                                                                              | Using Plunger dial gauge with Comparator Stand by Direct Method              | 0 to 100 µm                                                                       | 5.6 µm                                           |
| 169  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Height Gauge (L.C: 0.01 mm)                                                                                                 | Using Caliper Checker & Slip gauge Block by Comparison method                | 0 to 300 mm                                                                       | 7.93 µm                                          |
| 170  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Height Gauge (L.C: 0.01 mm)                                                                                                 | Using Caliper Checker, Slip gauge Block & Surface Plate by Comparison method | 0 to 600 mm                                                                       | 9.55 µm                                          |
| 171  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Inside Caliper (L.C: 0.01 mm)                                                                                               | Using Caliper Checker by Comparison method                                   | 0 to 300 mm                                                                       | 7.3 µm                                           |
| 172  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Le - Chatelier Mould - Diameter                                                                                             | Using Vernier Caliper by Comparison Method as per IS 5514                    | 20 mm to 50 mm                                                                    | 17.0 µm                                          |
| 173  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Le - Chatelier Mould - Height                                                                                               | Using Vernier Caliper by Comparison Method as per IS 5514                    | 20 mm to 50 mm                                                                    | 107.0 µm                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

28 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                 | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 174  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Le - Chatelier Mould - Split                                                                                                | Using Vernier Caliper by Comparison Method as per IS 5514                                      | 0.1 mm to 2 mm                                                                    | 17 µm                                            |
| 175  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Lever type Dial Gauge                                                                                                       | Using Dial Calibration Tester By Comparison method                                             | 0 to 0.8 mm                                                                       | 3.1 µm                                           |
| 176  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Lever type Dial Gauge (L.C. 0.001 mm)                                                                                       | Using Dial Calibration Tester/ Slip gauge set and comparator stand by comparison method        | 0 to 0.14 mm                                                                      | 3.3 µm                                           |
| 177  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Magnetic V Block (Parallelism)                                                                                              | Using Slip Gauge Block, Dial Test Indicator, Test Mandrel & surface Plate by Comparison method | Up to 150 mm x 100 mm x 7 mm                                                      | 3.85 µm                                          |
| 178  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Magnetic V Block (Squareness)                                                                                               | Using Master Cylinder, Slip Gauge Block & Surface Plate by Comparison method                   | Up to 150 mm x 100 mm x 7 mm                                                      | 7.4 µm                                           |
| 179  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Magnetic V Block (Symmetricity)                                                                                             | Using Slip Gauge Block, Dial Test Indicator, Test Mandrel & Surface Plate by Comparison method | Up to 150 mm x 100 mm x 7 mm                                                      | 5.2 µm                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

29 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                            | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 180  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Measuring Tape/ Circumference Tape/pi Tape/Woven metallic & glass fiber tape                                                | Using Measuring Scale & Tap calibration system by Comparison method                       | 0 to 50 m                                                                         | 42* $\sqrt{L}$ $\mu$ m, where L in m             |
| 181  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Micrometer Setting Rod                                                                                                      | Using Slip gauge set, Electronic probe & Comparator stand by Comparison method            | 25 mm to 100 mm                                                                   | 2.6 $\mu$ m                                      |
| 182  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Needle / Plunger of Vicat Apparatus - Diameter / Length                                                                     | Using Vernier Caliper by Comparison Method as per IS 5513                                 | 0 to 60 mm                                                                        | 10.9 $\mu$ m                                     |
| 183  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Penetrometer (L.C. : 0.1 mm)                                                                                                | Using Slip Gauge set by Direct Method                                                     | 0 to 40 mm                                                                        | 577.4 $\mu$ m                                    |
| 184  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Pistol caliper                                                                                                              | Using Slip Gauge Set '0' Grade by comparison method                                       | 1 mm to 150 mm                                                                    | 57.8 $\mu$ m                                     |
| 185  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Plug Gauge/Pin Gauge/ Air Plug gauge ,OD Gauge, Ball Diameter.                                                        | Using Gauge Block Set,Electronic probe with DRO and comparator Stand by Comparison Method | 1 mm to 150 mm                                                                    | 3.3 $\mu$ m                                      |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

30 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                                                   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 186  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plunger Dial Gauge (L.C: 0.001 mm)                                                                                          | Using Slip Gauge Block & Comparator stand by Comparison method                                                                   | 0 to 1 mm                                                                         | 2.5 µm                                           |
| 187  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plunger Dial Gauge (L.C: 0.01 mm)                                                                                           | Using Slip gauge set & Comparator stand by Comparison method                                                                     | 0 to 25 mm                                                                        | 3.2 µm                                           |
| 188  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plunger type dial gauge (L.C: 0.01 mm)                                                                                      | Using Dial Calibration tester by Comparison method                                                                               | 0 to 10 mm                                                                        | 3.3 µm                                           |
| 189  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Radius gauge / Radius measurement                                                                                           | using Profile projector by comparison method                                                                                     | 0.5 mm to 150 mm                                                                  | 5.2 µm                                           |
| 190  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Receiving Gauge/ Profile Gauge / Profile of Work Piece-Angle                                                                | Using Profile Projector by Direct Method                                                                                         | 0 to 360 degree                                                                   | 5 sec                                            |
| 191  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Receiving Gauge/ Profile Gauge / Profile of Work Piece-Length                                                               | Profile Projector/Gauge Block/Digital Caliper/Plunger dial gauge/lever dial gauge/Comparator Stand by Direct & Comparison Method | 0 to 300 mm                                                                       | 9.5 µm                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

31 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 192  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Receiving Gauge/ Profile Gauge / Profile of Work Piece-Radius                                                               | Using Profile Projector by Direct Method                                                      | 2.5 mm to 50 mm                                                                   | 4.6 µm                                           |
| 193  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Sand Pouring Cylinder - Diameter                                                                                            | Using Vernier Caliper by Comparison Method                                                    | 0 to 250 mm                                                                       | 16.3 µm                                          |
| 194  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Sand Pouring Cylinder - Height                                                                                              | Using Vernier Caliper by Comparison Method as per IS 2720-28                                  | 0 to 625 mm                                                                       | 18.7 µm                                          |
| 195  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Sand Replacement Cylinder/ Temping Rod/ Pocking Rod (dimension Measurement) (Diameter/Length/ measurement of Part)          | Using Dig. Caliper/Height Gauge/Depth Gauge/Dial gauge with comparator stand by Direct Method | 0 to 600 mm                                                                       | 37.4 µm                                          |
| 196  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Sine Bar (Centre Distance)                                                                                                  | Using height gauge and dial indicator direct Comparison                                       | upto 300 mm                                                                       | 4 µm                                             |
| 197  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Sine Bar (Flatness)                                                                                                         | Using electronic probe & surface plate by comparison method                                   | upto 300 mm                                                                       | 2.76 µm                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

32 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                           | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 198  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Sine Bar (Parallelism)                                                                                                      | Using electronic probe & surface plate by comparison method                              | upto 300 mm                                                                       | 3.4 µm                                           |
| 199  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Slump Cone - Diameter (Top / Bottom)                                                                                        | Using Vernier Caliper by Comparison Method                                               | 90 mm to 210 mm                                                                   | 24.0 µm                                          |
| 200  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Slump Cone - Height                                                                                                         | Using Vernier Caliper by Comparison Method                                               | 0 to 310 mm                                                                       | 18.7 µm                                          |
| 201  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Snap Gauge/ Width gauge                                                                                                     | Using Slip Gauge Block by Comparison method                                              | 1 mm to 150 mm                                                                    | 2.5 µm                                           |
| 202  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Standard Pin Gauge                                                                                                          | Using Slip Gauge Set '0' Grade, Electronic probe & Comparator Stand by Comparison method | 0.5 mm to 20 mm                                                                   | 1.2 µm                                           |
| 203  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Std wire gauge                                                                                                              | using profile projector by comparison method                                             | upto 25 mm                                                                        | 5.2 µm                                           |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

33 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                       | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 204  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Steel Scale                                                                                                                 | Using Measuring Scale & Tape calibration system by Comparison method | 0 to 1000 mm                                                                      | 27.66 µm                                         |
| 205  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Step Gauge                                                                                                                  | Using Digital Micrometer/ Profile Projector by Direct Method         | 0 to 25 mm                                                                        | 2.9 µm                                           |
| 206  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Straight edge                                                                                                               | Using height gauge & dial indicator by comparison method             | Upto 600 mm                                                                       | 4.24 µm                                          |
| 207  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Surface Plate                                                                                                               | Using Precision Spirit level by Comparison method                    | 3000 mm to 3000 mm                                                                | {1.87* $\sqrt{L+W}$ }/20 }µm/m                   |
| 208  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Taper Scale                                                                                                                 | Using Profile projector by Comparison Method                         | 1 mm to 50 mm                                                                     | 7.4 µm                                           |
| 209  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Test Sieve (Apparatus Size)                                                                                                 | Using Digital Caliper by Comparison method                           | 4 mm to 150 mm                                                                    | 14 µm                                            |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

34 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                           | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 210  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Test Sieves (Aperture Size)                                                                                                 | Using profile projector by comparison method                                             | 25 µm to 4 mm                                                                     | 4.5                                              |
| 211  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Thread Pitch Gauge - Angle                                                                                                  | Using profile projector by comparison method                                             | 55° & 60°                                                                         | 43.16"                                           |
| 212  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Thread Pitch Gauge - Pitch                                                                                                  | Using profile projector by comparison method                                             | 0.1 mm to 7.0 mm                                                                  | 5.2 µm                                           |
| 213  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Three Wire Set                                                                                                              | Using Slip Gauge Set '0' Grade, Electronic probe & Comparator Stand by Comparison method | 0.5 mm to 4 mm                                                                    | 1.76 µm                                          |
| 214  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Ultrasonic thickness gauge                                                                                                  | Using Slip gauge set grade "0" and Long gauge block by comparison method                 | Up to 200 mm                                                                      | 70.2 µm                                          |
| 215  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Vernier Caliper (Dial/Manual, Digital) (L.C: 0.01 mm)                                                                       | Using Caliper Checker, Slip gauge Block by Comparison method                             | 0 to 300 mm                                                                       | 8.0 µm                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

35 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 216  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Vernier Caliper - Dial/ Analogue/ Digital (L.C: 0.01 mm)                                                                    | Using Caliper Checker, Slip gauge Block by Comparison method                 | 0 to 600 mm                                                                       | 10.5 µm                                          |
| 217  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Vicat Test Apparatus Plunger Diameter (Needle)                                                                              | Using Dig. Micrometer by Direct Method                                       | 0.5 mm to 10 mm                                                                   | 3.0 µm                                           |
| 218  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)                  | Dial Calibration tester (L.C: 0.0002 mm)                                                                                    | Using Slip Gauge Block & Electronic probe by Comparison method               | 0 to 25 mm                                                                        | 0.96 µm                                          |
| 219  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)                  | Electronic Probe / LVDT (L.C: 0.0001 mm)                                                                                    | Using Slip Gauge Set '0' Grade & Comparator Stand by Comparison method       | 0 to 25 mm                                                                        | 2.34 µm                                          |
| 220  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)                  | Height Gauge (L.C. 0.001 mm)                                                                                                | Using Caliper Checker, Slip Gauge set grade "0" by Comparison method         | 0 to 600 mm                                                                       | 6 µm                                             |
| 221  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)                  | Height Gauge (L.C: 0.001 mm)                                                                                                | Using Caliper Checker, Slip gauge Block & Surface Plate by Comparison method | 0 to 300 mm                                                                       | 5.54 µm                                          |
| 222  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)                  | Microscope (Magnification)                                                                                                  | Using Glass Scale, Gauge Block by Comparison method                          | 1 X to 100 X                                                                      | 0.9 %                                            |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

36 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                           | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 223  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Projector X/Y Axis Movement (L.C: 0.001 mm)                                                                         | Using Glass scale /Slip Gauge Set '0' Grade by Comparison method                                    | 0 to 200 mm                                                                       | 2.0 µm                                           |
| 224  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Projector - Angle Scale (L.C: 1 s)                                                                                  | Using Angle Gauge Set by Comparison method                                                          | 0° to 360°                                                                        | 10 arc                                           |
| 225  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Projector-Magnification                                                                                             | Using Slip Gauge Set '0' Grade & Digital Caliper by Comparison Method                               | 10 X to 100 X                                                                     | 0.06 %                                           |
| 226  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Surface Roughness Tester (Portable) - Ra, Rz (L.C.: 0.001 µm)                                                               | Using Roughness Specimen Set by comparison method                                                   | 0 to 360 µm                                                                       | 6.8 %                                            |
| 227  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Universal length measuring machine (L.C: 0.0001 mm)                                                                         | Using Slip Gauge Set '0' Grade by Comparison method                                                 | 0 to 100 mm                                                                       | 0.7 µm                                           |
| 228  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Video Measuring Machine (Angular) (L.C: 1 arc second)                                                                       | Using Angle Gauge Set by Comparison method                                                          | 0° to 360°                                                                        | 10 s of arc                                      |
| 229  | MECHANICAL-DUROMETER                         | Durometer/Rubber Hardness Tester (Shore- A, B, E, O)                                                                        | Using Load Cell with Indicator & Fixture Based on ISO18898 & ASTM D 2240 on ISO 18898 & ASTM D 2240 | 0 to 100 Shore A                                                                  | 0.6 Shore A                                      |
| 230  | MECHANICAL-DUROMETER                         | Durometer/Rubber Hardness Tester (Shore- C, D, DO)                                                                          | Using Load Cell with Indicator & Fixture Based ISO 18898 & ASTM D 2240                              | 0 to 100 Shore                                                                    | 0.8 Shore D                                      |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

37 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                     | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                       | Calibration or Measurement Method or procedure                                                                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 231  | MECHANICAL-PRESSURE INDICATING DEVICES | Barometer / Barometric Pressure Gauge/Transmitter with Indicator, Transducers, Pressure Switch                                                    | Using digital barometer & pump,Digital Precision Multimeter by comparison method                                | 300 mbar to 1050 mbar                                                             | 1.52 mbar                                        |
| 232  | MECHANICAL-PRESSURE INDICATING DEVICES | Hydraulic Pressure Gauge ( Analog / Digital),Transmitter with Indicator, Transducers, Pressure Switch                                             | Using Digital pressure gauge, pressure pump & 6½ Digital multimeter based on DKD-R 6-1                          | 0 to 70 bar                                                                       | 0.12 bar                                         |
| 233  | MECHANICAL-PRESSURE INDICATING DEVICES | Hydraulic Pressure Gauge ( Analog / Digital),Transmitter with Indicator, Transducers, Pressure Switch                                             | Using Digital pressure gauge, pressure pump & 6½ Digital multimeter based on DKD-R 6-1                          | 0 to 700 bar                                                                      | 0.20 bar                                         |
| 234  | MECHANICAL-PRESSURE INDICATING DEVICES | Pneumatic Pressure Gauge ( Analog / Digital) , Transmitter with Indicator ,Transducers ,Pressure Switch                                           | Using Digital pressure gauge, pressure pump & 6½ Digital multimeter based on DKD-R 6-1 guidelines               | 0 bar to 40 bar                                                                   | 0.068 bar                                        |
| 235  | MECHANICAL-PRESSURE INDICATING DEVICES | Pressure -Pneumatic (Magnehelic Gauge, Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Transmitter, Pressure Switch, Manometer) | Using Digital Pressure Calibrator with Precision Low Pressure pump & 6½ Digital multimeter by Comparison method | (-) 195 mbar to 195 mbar                                                          | 0.23 mbar                                        |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

38 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                     | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                             | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 236  | MECHANICAL-PRESSURE INDICATING DEVICES | Vacuum Gauge ( Analog / Digital ) ,Transmitter with Indicator, Transducers, Vacuum Switch                                   | Using Digital Pressure Gauge, pressure pump & 6½ Digital multimeter based on DKD R6-1      | (-) 0.9 bar to 0 bar                                                              | 0.006 bar                                        |
| 237  | MECHANICAL-VOLUME                      | Glassware (Measuring Cylinder/ Jar, Volumetric Flask, Beaker, Glass Pipette, Dispenser, Burette)                            | Using Digital weighing balance (resolution: 0.01 g) by Gravimetric method as per ISO 4787  | > 1 l to 2 l                                                                      | 0.32 ml                                          |
| 238  | MECHANICAL-VOLUME                      | Glassware (Measuring Cylinder/ Jar, Volumetric Flask, Beaker, Glass Pipette, Dispenser, Burette)                            | Using Digital weighing balance (resolution: 0.001 g) by Gravimetric method as per ISO 4787 | >100 ml to 500 ml                                                                 | 9.8 µl                                           |
| 239  | MECHANICAL-VOLUME                      | Glassware (Measuring Cylinder/ Jar, Volumetric Flask, Beaker, Glass Pipette, Dispenser, Burette)                            | Using Digital weighing balance (resolution: 0.01 g) by Gravimetric method as per ISO 4787  | >2 l to 5 l                                                                       | 0.6 ml                                           |
| 240  | MECHANICAL-VOLUME                      | Glassware (Measuring Cylinder/ Jar, Volumetric Flask, Beaker, Glass Pipette, Dispenser, Burette)                            | Using Digital weighing balance (resolution: 0.01 mg) by Gravimetric method as per ISO 4787 | 1 ml to 100 ml                                                                    | 1.06 µl                                          |
| 241  | MECHANICAL-VOLUME                      | Micro pipette / Pipette                                                                                                     | Using Micro balance (resolution: 0.01 mg) as per ISO 8655: 2022                            | >100 µl to 1000 µl                                                                | 0.88 µl                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

39 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                    | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 242  | MECHANICAL-VOLUME                     | Micro pipette / Pipette                                                                                                     | Using Micro balance (resolution: 0.01 mg) as per ISO 8655: 2022 | >20 µl to 100 µl                                                                  | 0.20 µl                                          |
| 243  | MECHANICAL-VOLUME                     | Micro pipette / Pipette                                                                                                     | Using Micro balance (resolution: 0.01 mg) as per ISO 8655: 2022 | 1 ml to 10 ml                                                                     | 2.4 µl                                           |
| 244  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class - I & coarser) (Resolution= 0.0001 mg or coarser)                                                   | Using E1 class standard weights as per OIML R76-1               | 0 to 2.2 g                                                                        | 0.006 mg                                         |
| 245  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class I & coarser) (Readability 0.001 mg & Coarser)                                                       | Using E1 class standard weights as per OIML R76-1               | 0 to 22 g                                                                         | 0.011 mg                                         |
| 246  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class I & coarser) (Readability=0.01 mg & Coarser)                                                        | Using E1 class standard weights as per OIML R76-1               | 0 g to 220 g                                                                      | 0.08 mg                                          |
| 247  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class II & coarser) (Readability=0.1 g & Coarser)                                                         | Using E1 & F1 class standard weights as per OIML R76-1          | 0 to 6 kg                                                                         | 0.1 g                                            |
| 248  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class III & coarser) (Readability=0.5 g & Coarser)                                                        | Using E1 & F1 class standard weights as per OIML R76-1          | 0 to 30 kg                                                                        | 0.5 g                                            |
| 249  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class III & coarser) (Readability=1 g & Coarser)                                                          | Using F1 class standard weights as per OIML R76-1               | 0 to 100 kg                                                                       | 1 g                                              |
| 250  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class III) (Readability=5 g & Coarser)                                                                    | Using F1 class standard weights as per OIML R76-1               | 0 to 300 kg                                                                       | 5.9 g                                            |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

40 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                    | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                                          | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 251  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance 0 to 1Kg, (Class II & coarser) (Readability=0.01 g & Coarser)                                              | Using E1 & F1 class standard weights as per OIML R76-1                                                                  | 0 to 1 kg                                                                         | 0.01 g                                           |
| 252  | MECHANICAL-WEIGHTS                    | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004  | 1 g                                                                               | 0.010 mg                                         |
| 253  | MECHANICAL-WEIGHTS                    | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004  | 10 g                                                                              | 0.012 mg                                         |
| 254  | MECHANICAL-WEIGHTS                    | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01 mg) by Substitution method as per OIML R-111:2004 | 10 mg                                                                             | 0.0097 mg                                        |
| 255  | MECHANICAL-WEIGHTS                    | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004  | 100 g                                                                             | 0.022 mg                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

41 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                                         | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 256  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004 | 100 mg                                                                            | 0.0088 mg                                        |
| 257  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004 | 2 g                                                                               | 0.010 mg                                         |
| 258  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004 | 2 mg                                                                              | 0.0087 mg                                        |
| 259  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004 | 20 g                                                                              | 0.014 mg                                         |
| 260  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004 | 20 mg                                                                             | 0.009 mg                                         |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

42 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                                           | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|-----------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 261  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004   | 200 g                                                                             | 0.044 mg                                         |
| 262  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004   | 200 mg                                                                            | 0.01 mg                                          |
| 263  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004   | 5 g                                                                               | 0.011 mg                                         |
| 264  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution: 0.01 mg) by Substitution method as per OIML R-111:2004 | 5 mg                                                                              | 0.0096 mg                                        |
| 265  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004   | 50 g                                                                              | 0.017 mg                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

43 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                                              | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 266  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004      | 50 mg                                                                             | 0.0096 mg                                        |
| 267  | MECHANICAL-WEIGHTS | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004      | 500 mg                                                                            | 0.010 mg                                         |
| 268  | MECHANICAL-WEIGHTS | Accuracy class F2 & coarser                                                                                                 | Using F1 Class standard Weight & Digital Weighing Balance (resolution 0.001g) by Substitution method as per OIML R-111:2004 | 1 kg                                                                              | 2.3 mg                                           |
| 269  | MECHANICAL-WEIGHTS | Accuracy class F2 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004      | 1 mg                                                                              | 0.009 mg                                         |
| 270  | MECHANICAL-WEIGHTS | Accuracy class F2 & coarser                                                                                                 | Using F1 Class standard Weight & Digital Weighing Balance (resolution 0.01g) by Substitution method as per OIML R-111:2004  | 2 kg                                                                              | 9 mg                                             |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

44 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group               | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                                              | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 271  | MECHANICAL-WEIGHTS               | Accuracy class F2 & coarser                                                                                                 | Using F1 Class standard Weight & Digital Weighing Balance (resolution 0.01g) by Substitution method as per OIML R-111:2004  | 5 kg                                                                              | 13 mg                                            |
| 272  | MECHANICAL-WEIGHTS               | Accuracy class F2 & coarser                                                                                                 | Using F1 Class standard Weight & Digital Weighing Balance (resolution 0.001g) by Substitution method as per OIML R-111:2004 | 500 g                                                                             | 1.5 mg                                           |
| 273  | MECHANICAL-WEIGHTS               | Accuracy class M1 & coarser                                                                                                 | Using F1 Class standard Weight & Digital Weighing Balance (resolution 0.1g) by substitution method as per OIML R-111:2004   | 10 kg                                                                             | 84 mg                                            |
| 274  | MECHANICAL-WEIGHTS               | Accuracy class M1 & coarser                                                                                                 | Using F1 Class standard Weight & Digital Weighing Balance (resolution 0.1g) by Substitution method as per OIML R-111:2004   | 20 kg                                                                             | 88 mg                                            |
| 275  | THERMAL-SPECIFIC HEAT & HUMIDITY | Digital & Analog ThermoHygrometer, RH Sensors / Transmitters with controller/ Indicator/ Recorder/ Data Logger @ 25 °C      | Using RH &Temp sensor with indicator& humidity chamber by comparison method                                                 | 10 %Rh to 95 %Rh                                                                  | 1.09 %Rh                                         |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

45 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group               | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                              | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 276  | THERMAL-SPECIFIC HEAT & HUMIDITY | Digital & Analog ThermoHygrometer, RH Sensors / Transmitters with controller/Indicator/Recorder/Data Logger @ 50% Rh        | Using RH &Temp sensor with indicator& humidity chamber by comparison method | 10 °C to 50 °C                                                                    | 0.46 °C                                          |
| 277  | THERMAL-SPECIFIC HEAT & HUMIDITY | Indicator of Environment Chamber / Humidity Chamber/Humidity Source (Humidity Calibrator / Generator Chamber) @ 25 °C       | Using RH & Temp sensor with indicator by comparison method                  | 10 %Rh to 95 %Rh                                                                  | 1.07 %Rh                                         |
| 278  | THERMAL-TEMPERATURE              | Glass Thermometer                                                                                                           | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method   | (-) 80 °C to 50 °C                                                                | 0.13 °C                                          |
| 279  | THERMAL-TEMPERATURE              | Glass Thermometer                                                                                                           | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method   | 50 °C to 250 °C                                                                   | 0.13 °C                                          |
| 280  | THERMAL-TEMPERATURE              | Indicator with sensor of Autoclave (For Non-Medical Purpose Only - Single Position calibration)                             | Using PRT With super DAQ by Comparison Method                               | 120 °C to 138 °C                                                                  | 0.34 °C                                          |
| 281  | THERMAL-TEMPERATURE              | Indicator with sensor of Blackbody Source/ IR Thermal Sources (Single Position) @ emissivity 0.95                           | Using Infrared Pyrometer by comparison method                               | (-) 15 °C to 50 °C                                                                | 1.79 °C                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

46 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                       | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|-----------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 282  | THERMAL-TEMPERATURE | Indicator with sensor of Blackbody Source/ IR Thermal Sources (Single Position) @ emissivity 0.95                           | Using Infrared Pyrometer by comparison method                                        | 50 °C to 500 °C                                                                   | 2.1 °C                                           |
| 283  | THERMAL-TEMPERATURE | Indicator with sensor of Incubator, BOD Incubator (For Non Medical Purpose Only) (Single Position calibration)              | Using PRT With super DAQ by Comparison Method                                        | 20 °C to 60 °C                                                                    | 0.34 °C                                          |
| 284  | THERMAL-TEMPERATURE | IR Thermometer/IR Gun/RadiationPyrometer/ IRDetector/ ThermalImagers/Laserpointed/IRPyrometer @emmissivity 0.95             | Using Infrared Pyrometer & Black Body (Emissivity 0.95) Furnace by Comparison method | (-) 15 °C to 50 °C                                                                | 1.42 °C                                          |
| 285  | THERMAL-TEMPERATURE | IR Thermometer/IR Gun/RadiationPyrometer/ IRDetector/ ThermalImagers/Laserpointed/IRPyrometer @emmissivity 0.95             | Using Infrared Pyrometer & Black Body (Emissivity 0.95) Furnace by comparison method | 50 °C to 500 °C                                                                   | 1.82 °C                                          |
| 286  | THERMAL-TEMPERATURE | Temperature Indicator With sensor of muffle furnace, Dry Block Furnaces                                                     | Using PRT With super DAQ by Comparison Method (Single Position calibration)          | 250 °C to 650 °C                                                                  | 0.15 °C                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

47 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                   | Calibration or Measurement Method or procedure                                       | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)( $\pm$ ) |
|------|---------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
| 287  | THERMAL-TEMPERATURE | Temperature Indicators With sensor of (Freezers,Oven, Environment Chamber, Incubator, BOD Incubator, Liquid Bath/ Dry Block Furnaces) (Stability&Uniformity ) | Using PRT With super DAQ , by Comparison Method (Single Position calibration)        | 50 °C to 250 °C                                                                   | 0.15 °C                                                |
| 288  | THERMAL-TEMPERATURE | Temperature Indicators With sensor of (Freezers, Oven, Environment Chamber, Incubator, BOD Incubator, Liquid Bath/ Dry Block Furnaces, Metrology Well)        | Using PRT With super DAQ by Comparison Method (Single Position calibration)          | (-) 80 °C to 50 °C                                                                | 0.15 °C                                                |
| 289  | THERMAL-TEMPERATURE | Temperature Transmitter / RTD's/ Thermocouples /Temperature switch                                                                                            | Using PRT With super DAQ, Super DAQ and liquid Temperature Bath by Comparison Method | 50 °C to 250 °C                                                                   | 0.32 °C                                                |
| 290  | THERMAL-TEMPERATURE | Temperature Transmitter /RTD's/ Thermocouples /Temperature switch                                                                                             | Using PRT With super DAQ, Super DAQ and liquid Temperature Bath by Comparison Method | (-) 80 °C to 50 °C                                                                | 0.32 °C                                                |
| 291  | THERMAL-TEMPERATURE | Temperature Transmitter /RTD's/ Thermocouples /Temperature switch                                                                                             | Using PRT With super DAQ , Super DAQ and dry block calibrator, by Comparison Method  | 250 °C to 650 °C                                                                  | 0.35 °C                                                |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

48 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                   | Calibration or Measurement Method or procedure                            | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)( $\pm$ ) |
|------|---------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
| 292  | THERMAL-TEMPERATURE | Temperature Transmitter RTD's Thermocouples with (Controller/ Indicator/Data Logger/Recorder), Temperature Gauge, & Digital Thermometer, Temperature switch   | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method | 250 °C to 650 °C                                                                  | 0.15 °C                                                |
| 293  | THERMAL-TEMPERATURE | Temperature Transmitter RTD's/ Thermocouples with (Controller/ Indicator/Data Logger/Recorder), Temperature Gauge, & Digital Thermometer, Temperature switch  | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method | (-) 80 °C to 50 °C                                                                | 0.12 °C                                                |
| 294  | THERMAL-TEMPERATURE | Temperature Transmitter RTD's/ Thermocouples with (Controller/ Indicator/Data Logger/Recorder), Temperature Gauge, & Digital Thermometer, Temperature switch) | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method | 50 °C to 250 °C                                                                   | 0.13 °C                                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

49 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                           | Calibration or Measurement Method or procedure                                                                            | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 295  | THERMAL-TEMPERATURE | Temperature Transmitter, Thermocouples with (Controller /Indicator/Data Logger/ Recorder), Temperature Gauge, Digital Thermometer, temperature switch | Using S Type Thermocouple with Suoer DAQ and dry block calibrator by Comparison Method                                    | 650 °C to 1200 °C                                                                 | 1.43 °C                                          |
| 296  | THERMAL-TEMPERATURE | Temperature Transmitter/ Thermocouples /Temperature switch                                                                                            | Using S Type Thermocouple with Super DAQ and dry block calibrator by Comparison Method                                    | 650 °C to 1200 °C                                                                 | 1.55 °C                                          |
| 297  | THERMAL-TEMPERATURE | TemperatureIndicatorWith sensor of Thermal Sources ( Dry Block Furnaces, Metrology Well, Chambers, Dry Block calibrators, Muffle Furnace)             | Using Type-S Thermocouple S with Suoer DAQ By comparison method, (Singl e position calibration)                           | 600 to 1200                                                                       | 1.09                                             |
| 298  | THERMAL-TEMPERATURE | Thermal Sources of (Furnaces / Spatial Thermal Mapping (Multi-locations, Multiple Sensors)                                                            | Using Multiple N type thermocouple with Data Logger by comparison method (Multi-Position Calibration) - Minimum 9 sensors | 500 °C to 1200 °C                                                                 | 7.14 °C                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

50 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                      | Calibration or Measurement Method or procedure                                                                          | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|--------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 299  | THERMAL-TEMPERATURE | Thermal Sources of (Furnaces / Spatial Thermal Mapping (Multi-locations, Multiple Sensors, Temperature Uniformity Survey (TUS), Thermal Mapping) | Using Multiple N type thermocouple with Data Logger by comparison method (Multi-Position Calibration) - 9 Point sensors | 250 °C to 500 °C                                                                  | 5.03 °C                                          |
| 300  | THERMAL-TEMPERATURE | Thermal Sources of (Oven, Environmental Chambers / Furnaces / Heating Chambers / Incubators & BOD Incubators, Cold Room)                         | Using Multiple RTD Sensors with DataLogger by comparison method (Multi-Position Calibration) - Minimum 9 sensors        | (-) 80 °C to 50 °C                                                                | 2.0 °C                                           |
| 301  | THERMAL-TEMPERATURE | Thermal Sources of (Oven, Environmental Chambers / Furnaces / Heating Chambers / Incubators & BOD Incubators, Cold Room)                         | Using Multiple RTD Sensors with Data Logger by comparison method (Multi-Position Calibration) - Minimum 9 sensors       | 50 °C to 250 °C                                                                   | 2.50 °C                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

51 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No          | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---------------|-----------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| Site Facility |                                                           |                                                                                                                             |                                                                              |                                                                                   |                                                  |
| 1             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current @ 50 Hz                                                                                                          | Using 6½ Digital multimeter by Direct method                                 | 100 mA to 10 A                                                                    | 0.2 % to 0.3 %                                   |
| 2             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Current @ 50 Hz                                                                                                          | Using 6½ Digital multimeter by Direct method                                 | 30 µA to 100 mA                                                                   | 0.3 % to 0.2 %                                   |
| 3             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC High Current @ 50 Hz                                                                                                     | Using 6½ Digit Multimeter and Current Transformer (2000/5A) by Direct Method | 10 A to 2000 A                                                                    | 0.5 %                                            |
| 4             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC High voltage                                                                                                             | Using HV Probe with Digital Multimeter by Direct Method                      | 1 kV to 25 kV                                                                     | 5.4 %                                            |
| 5             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | AC Resistance @1 kHz                                                                                                        | Using LCR meter by direct method                                             | 1 ohm to 10 kohm                                                                  | 0.16 % to 0.11 %                                 |
| 6             | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | Capacitance @ 1 kHz                                                                                                         | Using LCR meter by direct method                                             | 10 pF to 1 µF                                                                     | 0.32 %                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

52 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                          | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 7    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | Capacitance @ 100 Hz                                                                                                        | Using LCR meter by direct method                                        | 10 $\mu$ F to 10 mF                                                               | 0.4 %                                            |
| 8    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | Energy - 3 Phase 4 wire (10 V to 440 V) (1A to 10A), Power Factor (0.1 to 1), (Lead/Lag)                                    | Using Three Phase Power Energy Logger / Calibrator by Comparison method | 0.3 kWh to 6 kWh                                                                  | 1.39 % to 1.43 %                                 |
| 9    | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | Energy - Single Phase, 2 wire (10 V to 440 V) (1 A to 10 A), Power Factor (0.1 to 1), (Lead/ Lag)                           | Using Three Phase Power Energy logger/ Calibrator by Comparison method  | 0.3 kWh to 6 kWh                                                                  | 1.39 % to 1.43 %                                 |
| 10   | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | Harmonic, Total Harmonic Distortion (THD) @ 50 Hz                                                                           | Using 3 Phase energy logger with Source by Direct / Comparison Method   | 30 V to 240 V (1 % to 10 %)                                                       | 0.7 %                                            |
| 11   | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure) | Inductance @ 100 Hz                                                                                                         | Using LCR meter by direct method                                        | 0.1 mH to 10 H                                                                    | 0.24 %                                           |
| 12   | ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)  | AC Current @ 50 Hz                                                                                                          | Using Multi Product Calibrator with current Coil by Direct method       | 20 A to 1000 A                                                                    | 0.4 % to 1.5 %                                   |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

53 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 13   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Current @ 50 Hz                                                                                                          | Using Multi Product Calibrator by Direct method                 | 3 mA to 20 A                                                                      | 0.12 % to 0.4 %                                  |
| 14   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Current @ 50 Hz                                                                                                          | Using Multi Product Calibrator by Direct method                 | 30 µA to 3 mA                                                                     | 0.55 % to 0.12 %                                 |
| 15   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Power (Active) 1 V to 1000V, 1 mA to 20A, UPF @ 50 Hz, 1 lead /Lag                                                       | Using Multi product Calibrator & Energy logger by Direct method | 1 W to 12000 W                                                                    | 0.30 % to 0.09 %                                 |
| 16   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Voltage (@ 50 Hz)                                                                                                        | Using Multi Product Calibrator by Direct method                 | 10 mV to 100 V                                                                    | 0.35 % to 0.04 %                                 |
| 17   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Voltage (@ 50 Hz)                                                                                                        | Using Multi Product Calibrator by Direct method                 | 100 V to 1000 V                                                                   | 0.04 % to 0.06 %                                 |
| 18   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | AC Voltage (50 HZ)                                                                                                          | Using Multi Product Calibrator by Direct method                 | 1 mV to 10 mV                                                                     | 2.5 % to 0.35 %                                  |
| 19   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Capacitance @ 1 kHz                                                                                                         | Using Multi Product Calibrator by Direct method                 | 0.5 nF to 10 nF                                                                   | 2.9 % to 0.58 %                                  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

54 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 20   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Capacitance @ 100 Hz                                                                                                        | Using Multi Product Calibrator by Direct method                 | 30 nF to 100 µF                                                                   | 0.5 % to 1.4 %                                   |
| 21   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Inductance @ 1 kHz                                                                                                          | Using LCR meter by direct method                                | 100 µH to 10 H                                                                    | 2.7 %                                            |
| 22   | ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source) | Power Factor (Lead/Lag) 240V/ 5A @ 50 Hz                                                                                    | Using Multi product Calibrator & Energy logger by Direct method | 0.1 PF to 1 PF                                                                    | 0.054 PF                                         |
| 23   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)               | DC Current                                                                                                                  | Using 6½ Digital multimeter by Direct method                    | 10 µA to 100 mA                                                                   | 0.7 % to 0.3 %                                   |
| 24   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)               | DC Current                                                                                                                  | Using 6½ Digital multimeter by Direct method                    | 100 mA to 10 A                                                                    | 0.3 %                                            |
| 25   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)               | DC High Current                                                                                                             | Using 6½ Digit Multimeter and Current Shunt by VI Method        | 10 A to 100 A                                                                     | 0.75 %                                           |
| 26   | ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)               | DC High Current                                                                                                             | Using 6½ Digit Multimeter and Current shunt by VI Method        | 100 A to 150 A                                                                    | 1.2 %                                            |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

55 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                         | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 27   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC High Voltage                                                                                                             | Using HV probe with DMM by direct method       | 1 kV to 35 kV                                                                     | 4.5 %                                            |
| 28   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC Voltage                                                                                                                  | Using 6½ Digital multimeter by Direct method   | 1 mV to 100 mV                                                                    | 0.4 % to 0.01 %                                  |
| 29   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC Voltage                                                                                                                  | Using 6½ Digital multimeter by Direct method   | 10 V to 1000 V                                                                    | 0.012 % to 0.007 %                               |
| 30   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | DC Voltage                                                                                                                  | Using 6½ Digital multimeter by Direct method   | 100 mV to 10 V                                                                    | 0.01 % to 0.012 %                                |
| 31   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance                                                                                                                  | Using Insulation Tester by direct method       | 1 Mohm to 1 Tohm                                                                  | 2.5 % to 2.9 %                                   |
| 32   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance                                                                                                                  | Using 6½ Digital multimeter by Direct method   | 1 Ohm to 1 Mohm                                                                   | 0.36 % to 0.013 %                                |
| 33   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance                                                                                                                  | Using 6½ Digital multimeter by Direct method   | 100 Mohm to 1000 Mohm                                                             | 0.96 % to 1.8 %                                  |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

56 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                         | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                   | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|--------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 34   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance - 4 wire                                                                                                         | Using micro ohm meter by direct method                           | 1 mohm to 100 mohm                                                                | 0.2 % to 0.05 %                                  |
| 35   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance - 4 wire                                                                                                         | Using microohm meter by direct method                            | 10 $\mu$ ohm to 1 mohm                                                            | 0.73 % to 0.32 %                                 |
| 36   | ELECTRO-TECHNICAL-DIRECT CURRENT (Measure) | Resistance - 4 Wire                                                                                                         | Using micro ohm meter by direct method                           | 100 mohm to 1 ohm                                                                 | 0.05 % to 0.09 %                                 |
| 37   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)  | DC Current                                                                                                                  | Using Multi Product Calibrator by Direct method                  | 1 $\mu$ A to 10 $\mu$ A                                                           | 2.50 % to 0.3 %                                  |
| 38   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)  | DC Current                                                                                                                  | Using Multi Product Calibrator by Direct method                  | 1 mA to 20 A                                                                      | 2.5 % to 0.4 %                                   |
| 39   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)  | DC Current                                                                                                                  | Using Multi Product Calibrator by Direct method                  | 10 $\mu$ A to 1 mA                                                                | 0.3 % to 2.5 %                                   |
| 40   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)  | DC Current                                                                                                                  | Using Multiproduct Calibrator with Current Coil by Direct method | 20 A to 1000 A                                                                    | 0.4 % to 0.13 %                                  |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

57 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 41   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Power (Active power) 10 V to 1000V, 0.1 A to 20A                                                                         | Using Multiproduct calibrator by direct method  | 100 W to 20000 W                                                                  | 0.97 %                                           |
| 42   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Voltage                                                                                                                  | Using Multi Product Calibrator by Direct method | 1 mV to 330 mV                                                                    | 0.4 % to 0.015 %                                 |
| 43   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | DC Voltage                                                                                                                  | Using Multi Product Calibrator by Direct method | 330 mV to 1000 V                                                                  | 0.015 % to 0.01 %                                |
| 44   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method | 1 Ohm to 30 Ohm                                                                   | 0.5 % to 0.07 %                                  |
| 45   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method | 30 kohm to 30 Mohm                                                                | 0.13 %                                           |
| 46   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method | 30 Mohm to 1000 Mohm                                                              | 0.13 % to 0.75 %                                 |
| 47   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire)                                                                                                         | Using Multi Product Calibrator by Direct method | 30 Ohm to 30 kohm                                                                 | 0.07 % to 0.13 %                                 |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

58 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 48   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method     | 1 Tohm                                                                            | 3.5 %                                            |
| 49   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method     | 10 Gohm                                                                           | 3.5 %                                            |
| 50   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method     | 10 Mohm                                                                           | 3.5 %                                            |
| 51   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method     | 100 Gohm                                                                          | 3.5 %                                            |
| 52   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method     | 100 Mohm                                                                          | 3.7 %                                            |
| 53   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 1000 V DC                                                                                             | Using High Resistance Box by Direct method     | 200 Gohm                                                                          | 3.5 %                                            |
| 54   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 1 Tohm                                                                            | 3.5 %                                            |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

59 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 55   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 10 Gohm                                                                           | 3.5 %                                            |
| 56   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 10 Mohm                                                                           | 3.5 %                                            |
| 57   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 100 Gohm                                                                          | 3.5 %                                            |
| 58   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 100 Mohm                                                                          | 3.5 %                                            |
| 59   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance (2 Wire) @ 5000 V DC                                                                                             | Using High Resistance Box by Direct method     | 200 Gohm                                                                          | 3.74 %                                           |
| 60   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 10 mohm                                                                           | 0.38 %                                           |
| 61   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source) | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method      | 10 Ohm                                                                            | 0.2 %                                            |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

60 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                              | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 62   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)       | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method                       | 100 mohm                                                                          | 0.2 %                                            |
| 63   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)       | Resistance - 4 wire                                                                                                         | Using Low Resistance Box by Direct method                       | 1000 mohm                                                                         | 0.2 %                                            |
| 64   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)       | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method                       | 1 mohm                                                                            | 0.55 %                                           |
| 65   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)       | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method                       | 100 Ohm                                                                           | 0.25 %                                           |
| 66   | ELECTRO-TECHNICAL-DIRECT CURRENT (Source)       | Resistance - 4 Wire                                                                                                         | Using Low Resistance Box by Direct method                       | 1000 Ohm                                                                          | 0.25 %                                           |
| 67   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | ORP Meter / ORP Indicator                                                                                                   | Using Multi product calibrator By Direct Method                 | (-) 1500 mV to 1500 mV                                                            | 0.5 %                                            |
| 68   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source) | Oscilloscope Amplitude @ 1 kHz Sinewave                                                                                     | Using Function / Arbitrary wave form generator by Direct method | 2 mVpp to 20 Vpp                                                                  | 0.45 % to 0.35 %                                 |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

61 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 69   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)    | Oscilloscope Amplitude @ 1 kHz Square wave                                                                                  | Using Function / Arbitrary wave form generator by Direct method | 2 mVpp to 20 Vpp                                                                  | 0.37 % to 0.35 %                                 |
| 70   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)    | Oscilloscope Bandwidth                                                                                                      | Using Function / Arbitrary wave form generator by Direct method | 50 kHz to 20 MHz                                                                  | 2.5 %                                            |
| 71   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)    | Oscilloscope Time Base                                                                                                      | Using Function / Arbitrary wave form generator by Direct method | 5 ns to 10 s                                                                      | 0.005 % to 0.01 %                                |
| 72   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)    | pH Meter (simulation method)                                                                                                | Using Multi product calibrator By Direct Method                 | 0 to 14 pH                                                                        | 0.13 %                                           |
| 73   | ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)    | TDS/ Conductivity Meter (Equivalent Resistance: 1 ohm to 2 Mohm)                                                            | Using Multiproduct calibrator by simulation method              | 0.5 $\mu$ S/cm to 1000 mS/cm                                                      | 1.72 % to 1.5 %                                  |
| 74   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | B Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method                 | 600 °C to 1800 °C                                                                 | 0.77 °C                                          |
| 75   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | E Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method                 | (-) 250 °C to 1000 °C                                                             | 0.63 °C                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD, HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Validity** 25/06/2025 to 24/06/2029

**Page No** 62 of 81

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                                 | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 76   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | J Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1200 °C                                                             | 0.40 °C                                          |
| 77   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | K Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1300 °C                                                             | 0.53 °C                                          |
| 78   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | N Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1300 °C                                                             | 0.54 °C                                          |
| 79   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | R Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 100 °C to 1700 °C                                                                 | 0.95 °C                                          |
| 80   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | RTD                                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 800 °C                                                              | 0.25 °C                                          |
| 81   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | S Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 100 °C to 1750 °C                                                                 | 0.65 °C                                          |
| 82   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure) | T Type Thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 400 °C                                                              | 0.74 °C                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

63 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)( $\pm$ ) |
|------|---------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
| 83   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | B Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 600 °C to 1800 °C                                                                 | 0.68 °C                                                |
| 84   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | E Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 250 °C to 1000 °C                                                             | 0.60 °C                                                |
| 85   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | J Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1050 °C                                                             | 0.74 °C                                                |
| 86   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | K Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1350 °C                                                             | 0.47 °C                                                |
| 87   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | N Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 1300 °C                                                             | 0.47 °C                                                |
| 88   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | RTD                                                                                                                         | Using Multi Product Calibrator by Direct method | (-) 200 °C to 800 °C                                                              | 0.3 °C                                                 |
| 89   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | S Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method | 0 °C to 1750 °C                                                                   | 0.66 °C                                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

64 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure               | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 90   | ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source) | T Type thermocouple                                                                                                         | Using Multi Product Calibrator by Direct method              | (-) 200 °C to 400 °C                                                              | 0.73 °C                                          |
| 91   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Frequency                                                                                                                   | Using 6½ Digital multimeter by Direct method                 | 10 Hz to 1 MHz                                                                    | 0.096 % to 0.02 %                                |
| 92   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Time                                                                                                                        | Using Digital Time Calibration standard by Comparison method | 0.1 s to 1 s                                                                      | 0.011 s                                          |
| 93   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Time                                                                                                                        | Using Digital Time Calibration standard by Comparison method | 1 s to 60 s                                                                       | 0.06 s                                           |
| 94   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Time                                                                                                                        | Using Digital Time Calibration standard by Comparison method | 3600 s to 43200 s                                                                 | 2.25 s                                           |
| 95   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Time                                                                                                                        | Using Digital Time Calibration standard by Comparison method | 43200 s to 86400 s                                                                | 2.50 s                                           |
| 96   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)      | Time                                                                                                                        | Using Digital Time Calibration standard by Comparison method | 60 s to 3600 s                                                                    | 2.25 s                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Validity** 25/06/2025 to 24/06/2029

**Page No** 65 of 81

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                          | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                          | Calibration or Measurement Method or procedure                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 97   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source) | Frequency                                                                                                                                            | Using Arbitrary Function generator by Direct method             | 1 MHz to 200 MHz                                                                  | 0.07 % to 0.5 %                                  |
| 98   | ELECTRO-TECHNICAL-TIME & FREQUENCY (Source) | Frequency                                                                                                                                            | Using Multi Product Calibrator by Direct method                 | 10 Hz to 2 MHz                                                                    | 0.5 % to 0.05 %                                  |
| 99   | FLUID FLOW-FLOW MEASURING DEVICES           | Volume flow rate(liquid flow meter & f low measuring equipments)                                                                                     | Using ultrasonic flow meter (medium water) by comparison method | 1 m <sup>3</sup> /hr to 1050 m <sup>3</sup> /hr                                   | 1.2 %                                            |
| 100  | MECHANICAL-ACCELERATION AND SPEED           | Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method              | 10 rpm to 100 rpm                                                                 | 0.46 rpm                                         |
| 101  | MECHANICAL-ACCELERATION AND SPEED           | Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method              | 10000 rpm to 30000 rpm                                                            | 6.01 rpm                                         |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Page No** 66 of 81

**Validity** 25/06/2025 to 24/06/2029

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                       | Measurand or Reference<br>Material/Type of instrument<br>or material to be calibrated<br>or measured / Quantity<br>Measured /Instrument | Calibration or Measurement<br>Method or procedure                            | Measurement range and<br>additional parameters<br>where applicable(Range<br>and Frequency) | * Calibration and<br>Measurement<br>Capability(CMC)(±) |
|------|------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------|--------------------------------------------------------|
| 102  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 10 rpm to 100 rpm                                                                          | 0.46 rpm                                               |
| 103  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 100 rpm to 1000<br>rpm                                                                     | 0.65 rpm                                               |
| 104  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 1000 rpm to 5000<br>rpm                                                                    | 2.02 rpm                                               |
| 105  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 10000 rpm to 30000<br>rpm                                                                  | 6.01 rpm                                               |
| 106  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 30000 rpm to 50000<br>rpm                                                                  | 6.0 rpm                                                |
| 107  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 5000 rpm to 10000<br>rpm                                                                   | 2.02 rpm                                               |
| 108  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 50000 rpm to 70000<br>rpm                                                                  | 7.01 rpm                                               |
| 109  | MECHANICAL-<br>ACCELERATION<br>AND SPEED | Digital Tachometer<br>(Non-contact type)                                                                                                | Using Standard<br>digital tachometer &<br>RPM source by<br>Comparison method | 70000 rpm to 90000<br>rpm                                                                  | 9.2 rpm                                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Validity** 25/06/2025 to 24/06/2029

**Page No** 67 of 81

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                       | Calibration or Measurement Method or procedure     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 110  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 1000 rpm to 5000 rpm                                                              | 2.02 rpm                                         |
| 111  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 30000 rpm to 50000 rpm                                                            | 6.0 rpm                                          |
| 112  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 5000 rpm to 10000 rpm                                                             | 2.02 rpm                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Page No** 68 of 81

**Validity** 25/06/2025 to 24/06/2029

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                       | Calibration or Measurement Method or procedure     | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 113  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 50000 rpm to 70000 rpm                                                            | 7.01 rpm                                         |
| 114  | MECHANICAL-ACCELERATION AND SPEED | Indicator of Centrifuge & Bitumen Extractor, Vibration machine, Abrasion testing machine, Washing machine, Tachometer calibrator, Laundrometer (Non-contact type) | Using Standard digital tachometer by Direct method | 70000 rpm to 90000 rpm                                                            | 9.2 rpm                                          |
| 115  | MECHANICAL-ACCELERATION AND SPEED | Indicator of RPM Indicating device of Centrifuge, Bitumen extractor, Vibrating Machine, Abrasion Tester (Non-contact type)                                        | Using Standard digital tachometer by direct method | 10 rpm to 100 rpm                                                                 | 0.46 rpm                                         |
| 116  | MECHANICAL-ACCELERATION AND SPEED | Indicator of RPM Indicating device of Centrifuge, Bitumen extractor, Vibrating Machine, Abrasion Tester (Non-contact type)                                        | Using Standard digital tachometer by direct method | 100 rpm to 1000 rpm                                                               | 0.65 rpm                                         |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

69 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------|-----------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 117  | MECHANICAL-ACCELERATION AND SPEED | Indicator of RPM Indicating device of Centrifuge, Bitumen extractor, Vibrating Machine, Abrasion Tester (Non-contact type)  | Using Standard digital tachometer by direct method                  | 10000 rpm to 30000 rpm                                                            | 6.01 rpm                                         |
| 118  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                   | Using Standard digital tachometer & RPM source by Comparison method | 10 rpm to 100 rpm                                                                 | 0.45 rpm                                         |
| 119  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                   | Using Standard digital tachometer & RPM source by Comparison method | 100 rpm to 1000 rpm                                                               | 0.82 rpm                                         |
| 120  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                   | Using Standard digital tachometer & RPM source by Comparison method | 1000 rpm to 3000 rpm                                                              | 2.12 rpm                                         |
| 121  | MECHANICAL-ACCELERATION AND SPEED | Tachometer (Contact type)                                                                                                   | Using Standard digital tachometer & RPM source by Comparison method | 3000 rpm to 5000 rpm                                                              | 2.2 rpm                                          |
| 122  | MECHANICAL-ACCELERATION AND SPEED | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                  | 10 rpm to 100 rpm                                                                 | 0.44 rpm                                         |
| 123  | MECHANICAL-ACCELERATION AND SPEED | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                  | 10 rpm to 100 rpm                                                                 | 0.45 rpm                                         |
| 124  | MECHANICAL-ACCELERATION AND SPEED | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                  | 100 rpm to 1000 rpm                                                               | 0.82 rpm                                         |
| 125  | MECHANICAL-ACCELERATION AND SPEED | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                  | 1000 rpm to 3000 rpm                                                              | 2.12 rpm                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

70 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                            | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                      | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 126  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer & RPM source by Comparison method | 3000 rpm to 5000 rpm                                                              | 2.2 rpm                                          |
| 127  | MECHANICAL-ACCELERATION AND SPEED                             | Tachometer calibrator, Trademill (Contact type)                                                                             | Using Standard digital tachometer by Direct method                  | 3000 rpm to 5000 rpm                                                              | 2.2 rpm                                          |
| 128  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bench Centre (coaxility)                                                                                                    | Using Test mandrel & Dial test indicator by Comparison method       | Up to 300 mm                                                                      | 8.0 µm                                           |
| 129  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bench Centre (parallelism)                                                                                                  | Using Test mandrel & Dial test indicator by Comparison method       | Up to 300 mm                                                                      | 8.0 µm                                           |
| 130  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Straight edge                                                                                                               | Using height gauge & dial indicator by comparison method            | Upto 600 mm                                                                       | 4.24 µm                                          |
| 131  | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Surface Plate                                                                                                               | Using Precision Spirit level by Comparison method                   | 3000 mm to 3000 mm                                                                | {1.87* $\sqrt{L+W}$ }/20 }µm/m                   |
| 132  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)                  | Microscope (Magnification)                                                                                                  | Using Glass Scale, Gauge Block by Comparison method                 | 1 X to 100 X                                                                      | 0.9 %                                            |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

71 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                           | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                          | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 133  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Projector X/Y Axis Movement (L.C: 0.001 mm)                                                                         | Using Glass scale /Slip Gauge Set '0' Grade by Comparison method        | 0 to 200 mm                                                                       | 2.0 µm                                           |
| 134  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Projector-Magnification                                                                                             | Using Slip Gauge Set '0' Grade & Digital Caliper by Comparison Method   | 10 X to 100 X                                                                     | 0.06 %                                           |
| 135  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Surface Roughness Tester (Portable) - Ra, Rz (L.C.: 0.001 µm)                                                               | Using Roughness Specimen Set by comparison method                       | 0 to 360 µm                                                                       | 6.8 %                                            |
| 136  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Universal length measuring machine (L.C: 0.0001 mm)                                                                         | Using Slip Gauge Set '0' Grade by Comparison method                     | 0 to 100 mm                                                                       | 0.7 µm                                           |
| 137  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Video Measuring Machine (Angular) (L.C: 1 arc second)                                                                       | Using Angle Gauge Set by Comparison method                              | 0° to 360°                                                                        | 10 s of arc                                      |
| 138  | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Video Measuring Machine - Linear (L.C: 0.0001 mm)                                                                           | Using Glass scale /Slip Gauge Set '0' Grade by Comparison method        | 0 to 200 mm                                                                       | 1.3 µm                                           |
| 139  | MECHANICAL-HARDNESS TESTING MACHINES         | Rockwell Hardness Testing Machine                                                                                           | Using Reference Blocks as per IS 1586 (Part 2): 2018 by Indirect Method | HRBW                                                                              | 0.9 HRBW                                         |
| 140  | MECHANICAL-HARDNESS TESTING MACHINES         | Rockwell Hardness Testing Machine                                                                                           | Using Reference Blocks as per IS 1586 (Part 2): 2018 by Indirect Method | HRC                                                                               | 0.9 HRC                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD, HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257

**Validity** 25/06/2025 to 24/06/2029

**Page No** 72 of 81

**Last Amended on** 18/07/2025

| S.No | Discipline / Group                     | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                       | Calibration or Measurement Method or procedure                                                                  | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 141  | MECHANICAL-PRESSURE INDICATING DEVICES | Barometer / Barometric Pressure Gauge/Transmitter with Indicator, Transducers, Pressure Switch                                                    | Using digital barometer & pump,Digital Precision Multimeter by comparison method                                | 300 mbar to 1050 mbar                                                             | 1.52 mbar                                        |
| 142  | MECHANICAL-PRESSURE INDICATING DEVICES | Hydraulic Pressure Gauge ( Analog / Digital),Transmitter with Indicator, Transducers, Pressure Switch                                             | Using Digital pressure gauge, pressure pump & 6½ Digital multimeter based on DKD-R 6-1                          | 0 to 70 bar                                                                       | 0.12 bar                                         |
| 143  | MECHANICAL-PRESSURE INDICATING DEVICES | Hydraulic Pressure Gauge ( Analog / Digital),Transmitter with Indicator, Transducers, Pressure Switch                                             | Using Digital pressure gauge, pressure pump & 6½ Digital multimeter based on DKD-R 6-1                          | 0 to 700 bar                                                                      | 0.20 bar                                         |
| 144  | MECHANICAL-PRESSURE INDICATING DEVICES | Pneumatic Pressure Gauge ( Analog / Digital) , Transmitter with Indicator ,Transducers ,Pressure Switch                                           | Using Digital pressure gauge, pressure pump & 6½ Digital multimeter based on DKD-R 6-1 guidelines               | 0 bar to 40 bar                                                                   | 0.068 bar                                        |
| 145  | MECHANICAL-PRESSURE INDICATING DEVICES | Pressure -Pneumatic (Magnehelic Gauge, Pressure Gauge, Pressure Indicator, Pressure Calibrator, Pressure Transmitter, Pressure Switch, Manometer) | Using Digital Pressure Calibrator with Precision Low Pressure pump & 6½ Digital multimeter by Comparison method | (-) 195 mbar to 195 mbar                                                          | 0.23 mbar                                        |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

73 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                     | Calibration or Measurement Method or procedure                                        | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 146  | MECHANICAL-PRESSURE INDICATING DEVICES                    | Vacuum Gauge ( Analog / Digital ) ,Transmitter with Indicator, Transducers, Vacuum Switch                                                       | Using Digital Pressure Gauge, pressure pump & 6½ Digital multimeter based on DKD R6-1 | (-) 0.9 bar to 0 bar                                                              | 0.006 bar                                        |
| 147  | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Universal testing Machine ,Compression testing machine ,Spring testing machine, Load testing machine, Fluxural testing machine Compression mode | Using Force Proving Instrument as per IS 1828(Part 1): 2022 / ISO 7500 - 1: 2018      | 10 kN to 100 kN                                                                   | 0.67 %                                           |
| 148  | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Universal testing Machine ,Compression testing machine ,Spring testing machine, Load testing machine, Fluxural testing machine Compression mode | Using Force Proving Instrument as per IS 1828(Part 1): 2022 / ISO 7500 - 1: 2018      | 100 kN to 1000 kN                                                                 | 0.69 %                                           |
| 149  | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Universal testing Machine ,Compression testing machine ,Spring testing machine, Load testing machine, Fluxural testing machine Compression mode | Using Force Proving Instrument as per IS 1828(Part 1): 2022 / ISO 7500 - 1: 2018      | 200 kN to 2000 kN                                                                 | 0.57 %                                           |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD, HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257 **Page No** 74 of 81

**Validity** 25/06/2025 to 24/06/2029 **Last Amended on** 18/07/2025

| S.No | Discipline / Group                                        | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                     | Calibration or Measurement Method or procedure                                                                                             | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|-----------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 150  | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Universal testing Machine ,Compression testing machine, Spring testing machine, Load testing machine, Fluxural testing machine Compression mode | Using Load cell with indicator as per IS 1828 (Part 1): 2022 / ISO 7500 - 1: 2018                                                          | 50 N to 10 kN                                                                     | 0.87 %                                           |
| 151  | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Universal testing Machine ,Tensile testing machine Spring testing machine, Load testing machine Tension Mode                                    | Using Using Load cell with indicator as per IS 1828 (Part 1): 2022 / ISO 7500 - 1: 2018 as per IS 1828 (Part 1): 2022 / ISO 7500 - 1: 2018 | 50 N to 10 kN                                                                     | 0.69 %                                           |
| 152  | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Verification/ Calibration of Displacement Measuring System and Device Used in Material Testing Machine (Encoder/ Actuator) - Class B            | Using Digital Height Gauge As per ASTM E 2309 by Comparison Method                                                                         | 0 to 600 mm                                                                       | 90 µm                                            |
| 153  | MECHANICAL-UTM, TENSION CREEP AND TORSION TESTING MACHINE | Verification/Calibration of Speed of Material Testing Machine                                                                                   | Using Digital Height Gauge, Lever Dial and Stop Watch as per ASTM E 2658 by Comparison Method                                              | 0 to 200 mm/min.                                                                  | 1.5 mm/min.                                      |
| 154  | MECHANICAL-WEIGHING SCALE AND BALANCE                     | Weighing Balance (Class - I & coarser) (Resolution= 0.0001 mg or coarser)                                                                       | Using E1 class standard weights as per OIML R76-1                                                                                          | 0 to 2.2 g                                                                        | 0.006 mg                                         |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

75 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group                    | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                                                         | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------------------------|-----------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 155  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class I & coarser) (Readability 0.001 mg & Coarser)                                                       | Using E1 class standard weights as per OIML R76-1                                                                      | 0 to 22 g                                                                         | 0.011 mg                                         |
| 156  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class I & coarser) (Readability=0.01 mg & Coarser)                                                        | Using E1 class standard weights as per OIML R76-1                                                                      | 0 g to 220 g                                                                      | 0.08 mg                                          |
| 157  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class II & coarser) (Readability=0.1 g & Coarser)                                                         | Using E1 & F1 class standard weights as per OIML R76-1                                                                 | 0 to 6 kg                                                                         | 0.1 g                                            |
| 158  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class III & coarser) (Readability=0.5 g & Coarser)                                                        | Using E1 & F1 class standard weights as per OIML R76-1                                                                 | 0 to 30 kg                                                                        | 0.5 g                                            |
| 159  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance (Class III & coarser) (Readability=1 g & Coarser)                                                          | Using F1 class standard weights as per OIML R76-1                                                                      | 0 to 100 kg                                                                       | 1 g                                              |
| 160  | MECHANICAL-WEIGHING SCALE AND BALANCE | Weighing Balance 0 to 1Kg, (Class II & coarser) (Readability=0.01 g & Coarser)                                              | Using E1 & F1 class standard weights as per OIML R76-1                                                                 | 0 to 1 kg                                                                         | 0.01 g                                           |
| 161  | MECHANICAL-WEIGHTS                    | Accuracy class F1 & coarser                                                                                                 | Using E1 class standard weights & semi micro balance (resolution 0.01mg) by Substitution method as per OIML R-111:2004 | 1 g                                                                               | 0.010 mg                                         |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

76 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group               | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                              | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|----------------------------------|-----------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 162  | THERMAL-SPECIFIC HEAT & HUMIDITY | Digital & Analog ThermoHygrometer, RH Sensors / Transmitters with controller/ Indicator/ Recorder/ Data Logger @ 25 °C      | Using RH &Temp sensor with indicator& humidity chamber by comparison method | 10 %Rh to 95 %Rh                                                                  | 1.09 %Rh                                         |
| 163  | THERMAL-SPECIFIC HEAT & HUMIDITY | Digital & Analog ThermoHygrometer, RH Sensors / Transmitters with controller/Indicator/R ecorder/Data Logger @ 50% Rh       | Using RH &Temp sensor with indicator& humidity chamber by comparison method | 10 °C to 50 °C                                                                    | 0.46 °C                                          |
| 164  | THERMAL-SPECIFIC HEAT & HUMIDITY | Indicator of Environment Chamber / Humidity Chamber/Humidity Source (Humidity Calibrator / Generator Chamber) @ 25 °C       | Using RH & Temp sensor with indicator by comparison method                  | 10 %Rh to 95 %Rh                                                                  | 1.07 %Rh                                         |
| 165  | THERMAL-TEMPERATURE              | Glass Thermometer                                                                                                           | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method   | (-) 80 °C to 50 °C                                                                | 0.13 °C                                          |
| 166  | THERMAL-TEMPERATURE              | Glass Thermometer                                                                                                           | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method   | 50 °C to 250 °C                                                                   | 0.13 °C                                          |
| 167  | THERMAL-TEMPERATURE              | Indicator with sensor of Autoclave (For Non-Medical Purpose Only - Single Position calibration)                             | Using PRT With super DAQ by Comparison Method                               | 120 °C to 138 °C                                                                  | 0.34 °C                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

77 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure                                       | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|-----------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 168  | THERMAL-TEMPERATURE | Indicator with sensor of Blackbody Source/ IR Thermal Sources (Single Position) @ emissivity 0.95                           | Using Infrared Pyrometer by comparison method                                        | (-) 15 °C to 50 °C                                                                | 1.79 °C                                          |
| 169  | THERMAL-TEMPERATURE | Indicator with sensor of Blackbody Source/ IR Thermal Sources (Single Position) @ emissivity 0.95                           | Using Infrared Pyrometer by comparison method                                        | 50 °C to 500 °C                                                                   | 2.1 °C                                           |
| 170  | THERMAL-TEMPERATURE | Indicator with sensor of Incubator, BOD Incubator (For Non Medical Purpose Only) (Single Position calibration)              | Using PRT With super DAQ by Comparison Method                                        | 20 °C to 60 °C                                                                    | 0.34 °C                                          |
| 171  | THERMAL-TEMPERATURE | IR Thermometer/IR Gun/RadiationPyrometer/ IRDetector/ ThermalImagers/Laserpointed/IRPyrometer @emmissivity 0.95             | Using Infrared Pyrometer & Black Body (Emissivity 0.95) Furnace by Comparison method | (-) 15 °C to 50 °C                                                                | 1.42 °C                                          |
| 172  | THERMAL-TEMPERATURE | IR Thermometer/IR Gun/RadiationPyrometer/ IRDetector/ ThermalImagers/Laserpointed/IRPyrometer @emmissivity 0.95             | Using Infrared Pyrometer & Black Body (Emissivity 0.95) Furnace by comparison method | 50 °C to 500 °C                                                                   | 1.82 °C                                          |
| 173  | THERMAL-TEMPERATURE | Temperature Indicator With sensor of muffle furnace, Dry Block Furnaces                                                     | Using PRT With super DAQ by Comparison Method (Single Position calibration)          | 250 °C to 650 °C                                                                  | 0.15 °C                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

78 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                   | Calibration or Measurement Method or procedure                                       | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 174  | THERMAL-TEMPERATURE | Temperature Indicators With sensor of (Freezers,Oven, Environment Chamber, Incubator, BOD Incubator, Liquid Bath/ Dry Block Furnaces) (Stability&Uniformity ) | Using PRT With super DAQ , by Comparison Method (Single Position calibration)        | 50 °C to 250 °C                                                                   | 0.15 °C                                          |
| 175  | THERMAL-TEMPERATURE | Temperature Indicators With sensor of (Freezers, Oven, Environment Chamber, Incubator, BOD Incubator, Liquid Bath/ Dry Block Furnaces, Metrology Well)        | Using PRT With super DAQ by Comparison Method (Single Position calibration)          | (-) 80 °C to 50 °C                                                                | 0.15 °C                                          |
| 176  | THERMAL-TEMPERATURE | Temperature Transmitter / RTD's/ Thermocouples /Temperature switch                                                                                            | Using PRT With super DAQ, Super DAQ and liquid Temperature Bath by Comparison Method | 50 °C to 250 °C                                                                   | 0.32 °C                                          |
| 177  | THERMAL-TEMPERATURE | Temperature Transmitter /RTD's/ Thermocouples /Temperature switch                                                                                             | Using PRT With super DAQ, Super DAQ and liquid Temperature Bath by Comparison Method | (-) 80 °C to 50 °C                                                                | 0.32 °C                                          |
| 178  | THERMAL-TEMPERATURE | Temperature Transmitter /RTD's/ Thermocouples /Temperature switch                                                                                             | Using PRT With super DAQ , Super DAQ and dry block calibrator, by Comparison Method  | 250 °C to 650 °C                                                                  | 0.35 °C                                          |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

79 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                                   | Calibration or Measurement Method or procedure                            | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)( $\pm$ ) |
|------|---------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
| 179  | THERMAL-TEMPERATURE | Temperature Transmitter RTD's Thermocouples with (Controller/ Indicator/Data Logger/Recorder), Temperature Gauge, & Digital Thermometer, Temperature switch   | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method | 250 °C to 650 °C                                                                  | 0.15 °C                                                |
| 180  | THERMAL-TEMPERATURE | Temperature Transmitter RTD's/ Thermocouples with (Controller/ Indicator/Data Logger/Recorder), Temperature Gauge, & Digital Thermometer, Temperature switch  | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method | (-) 80 °C to 50 °C                                                                | 0.12 °C                                                |
| 181  | THERMAL-TEMPERATURE | Temperature Transmitter RTD's/ Thermocouples with (Controller/ Indicator/Data Logger/Recorder), Temperature Gauge, & Digital Thermometer, Temperature switch) | Using PRT With super DAQ and liquid Temperature Bath by Comparison Method | 50 °C to 250 °C                                                                   | 0.13 °C                                                |



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-3257

**Page No**

80 of 81

**Validity**

25/06/2025 to 24/06/2029

**Last Amended on**

18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                           | Calibration or Measurement Method or procedure                                                                            | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 182  | THERMAL-TEMPERATURE | Temperature Transmitter, Thermocouples with (Controller /Indicator/Data Logger/ Recorder), Temperature Gauge, Digital Thermometer, temperature switch | Using S Type Thermocouple with Suoer DAQ and dry block calibrator by Comparison Method                                    | 650 °C to 1200 °C                                                                 | 1.43 °C                                          |
| 183  | THERMAL-TEMPERATURE | Temperature Transmitter/ Thermocouples /Temperature switch                                                                                            | Using S Type Thermocouple with Super DAQ and dry block calibrator by Comparison Method                                    | 650 °C to 1200 °C                                                                 | 1.55 °C                                          |
| 184  | THERMAL-TEMPERATURE | TemperatureIndicatorWith sensor of Thermal Sources ( Dry Block Furnaces, Metrology Well, Chambers, Dry Block calibrators, Muffle Furnace)             | Using Type-S Thermocouple S with Suoer DAQ By comparison method, (Singl e position calibration)                           | 600 to 1200                                                                       | 1.09                                             |
| 185  | THERMAL-TEMPERATURE | Thermal Sources of (Furnaces / Spatial Thermal Mapping (Multi-locations, Multiple Sensors)                                                            | Using Multiple N type thermocouple with Data Logger by comparison method (Multi-Position Calibration) - Minimum 9 sensors | 500 °C to 1200 °C                                                                 | 7.14 °C                                          |





# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** KL TECHNOLOGY SOLUTIONS, E-5, NEHRU GROUND, FARIDABAD,  
HARYANA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3257 **Page No** 81 of 81

**Validity** 25/06/2025 to 24/06/2029 **Last Amended on** 18/07/2025

| S.No | Discipline / Group  | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument                      | Calibration or Measurement Method or procedure                                                                          | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|------|---------------------|--------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------|
| 186  | THERMAL-TEMPERATURE | Thermal Sources of (Furnaces / Spatial Thermal Mapping (Multi-locations, Multiple Sensors, Temperature Uniformity Survey (TUS), Thermal Mapping) | Using Multiple N type thermocouple with Data Logger by comparison method (Multi-Position Calibration) - 9 Point sensors | 250 °C to 500 °C                                                                  | 5.03 °C                                          |
| 187  | THERMAL-TEMPERATURE | Thermal Sources of (Oven, Environmental Chambers / Furnaces / Heating Chambers / Incubators & BOD Incubators, Cold Room)                         | Using Multiple RTD Sensors with DataLogger by comparison method (Multi-Position Calibration) - Minimum 9 sensors        | (-) 80 °C to 50 °C                                                                | 2.0 °C                                           |
| 188  | THERMAL-TEMPERATURE | Thermal Sources of (Oven, Environmental Chambers / Furnaces / Heating Chambers / Incubators & BOD Incubators, Cold Room)                         | Using Multiple RTD Sensors with Data Logger by comparison method (Multi-Position Calibration) - Minimum 9 sensors       | 50 °C to 250 °C                                                                   | 2.50 °C                                          |

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.